

Heating Mantle (Digital, Magnetic)

User Manual



98-I-B Manual Heating Mantle
 98-II-B Magnetic Heating Mantle
 98-I-C/98-II-C Digital Heating Mantle

98-III-B Magnetic and Digital Heating Mantle

98-IV-B Rows Manual Heating Mantle 98-V-B Rows Magnetic Heating Mantle

Huanghua Faithful Instrument Co.,Ltd

No.7 Road, Economic & Technique Zone, Huanghua City, HebeiProvince, China

Tel: (86)317-533-7349 / 5327117 Fax: (86)317-533-8349

Website: www.faithful.cc

E-mail: info@labinstrument.cn info@.faithful.cc

Contents

Contents	1
Preface	2
Warranty	2
1 Safety Instructions	3
2 Proper Use	5
3 Inspection	5
3.1 Receiving Inspection	5
3.2 Listing of Items	5
4 Trial runs	6
5 Operating Modes	7
5.1 98-I-B Manual Heating Mantle and	
98-IV-B Rows Manual Heating Mantle	7
5.2 98-II-B Magnetic Heating Mantle and	
98-V-B Rows Magnetic Heating Mantle	7
5.3 98-I-C Digital Heating Mantle	8
5.4 98-II-C Digital Heating Mantle	8
5.5 98-III-B Magnetic and Digital Heating Mantle	9
6 Faults	9
7 Maintenance and Cleaning	10
8 Storage and transportation	10
9 Working condition	11
10 Main technical parameters	11

Preface

Thank you for purchasing our products: Heating Mantle. Users should read this Manual carefully, follow the instructions and procedures, and beware of all the cautions when using this instrument.

Warranty

You have purchased a Faithful instrument. This instrument is warranted to be free from defects in materials and workmanship under normal use and service, for a period of 12 months from the date of invoice. The warranty is extended only to the original purchaser. It shall not apply to any product or parts which have been damaged on account of improper installation, improper connections, misuse, accident or abnormal conditions of operation.

For claim under the warranty please contact with us. You many also send the instrument direct to our works or we send you the spare parts to help you resolve this problem in next order, enclosing the invoice copy and by giving reasons for the claim. You would be solely liable for freight costs.

1 Safety Instructions



Connect the device to an earthed power supply to ensure safety of machine and experiment; connect the power as the machine required.



This equipment is forbid to use in inflammable and explosive, poisonous and strong corrosive experiments.



Make sure horizontal installation.



Non-professionals are not allowed to disassemble and repair this machine.



Pay attention to the set temperature while dealing with the inflammable matters.



Make sure dry the resin container, if the temperature is setting too high by accident, the container would be dissolved and then fall on the heater to cause fire.



Overfilled of sample will lead to overheat of working room under parts, which will dissolve the inflammable material and cause fire.



While the machine is working, don't touch the top, window and exhaust port of the device to protect from high-temperature burns.



Read the instruction book before operation.

Table 1

- When working, ware the personal guard to avoid the risk from:
 - Splashing and evaporation of liquids
 - Release of toxic or combustible gases.
- Set up the instrument in a spacious area on a stable, clean, non-slip, dry and fireproof surface; do not operate the instrument in explosive atmospheres, with hazardous substances or under water.
- Temperature must always be set to at least 25°C lower than the fir point of the media used.
- Beware of hazards due to:
 - Flammable material or media with a low boiling temperature
 - Overfilling of media
 - Unsafe container
- · Process pathogenic materials only in closed vessels.
- Check the instrument and accessories before hand for damage each time you use them. Do not use damaged components. Safe operation is only guaranteed with the accessories described in the "Accessories" chapter. Accessories must be securely attached to the device and cannot come off by themselves. Always disconnect the plug before fitting accessories.
- Ensure that the external temperature sensor is inserted in the media to a depth of at least 20mm.
- When using metal vessels, do not place the temperature sensors on the bottom of the
 vessel. Placing sensors on the vessel bottom can cause excessively high temperature
 to be measured especially in media which have poor conductivity. The tip of the
 measuring sensor must be at least 5mm from the vessel bottom, a distance of 10mm
 is ideal.
- The instrument can only be disconnected from the main power supply by pulling out the mains plug or the connector plug.
- The voltage stated on the label must correspond to the main power supply.
- Ensure that the mains power supply cable does not touch the heating base plate. Do
 not cover the device.
- · Keep away from high magnetic field.

2 Proper Uses

The instrument is designed for mixing and / or heating liquids in schools, laboratories or factories. This device is not suitable for using in residential areas.

3 Inspections

3.1 Receiving Inspection

Unpack the equipment carefully and check for any damages which may have arisen during transport. If it happens, please contact manufacturer for technical support.

Note:

If there is any apparent damage to the system,

3.2 Listing of Items

The packing includes the following items:

Please do not plug it into the power line.

98-I-B Manual He	ating Mantle	98-II-B Magnetic Heatin	ng Mantle
Items	Qty	Items	Qty
Main Unit	11	Main Unit	<u> </u>
Power Cable	11	Power Cable	<u> </u>
User Manual	1_	User Manual	<u> </u>
Table2		Stir Bar	<u> </u>
		Table 3	
98-I-C/98-II-C Dig	gital Heating Mantle	98-III-B Magnetic	and Digital Heating
Mantle			
Items	Qty	Items	Qty
Main Unit	<u> </u>	Main Unit	<u> </u>
Power Cable	<u> </u>	Power Cable	<u> </u>
User Manual 1		User Manual	<u> </u>
Rack Rods 1		Rack Rods	<u> </u>
Table 4	4	Stirrer Bar	<u> </u>

98-IV-B	Rows	Manual	Heating	Mantle
00 II D		manaa	. ioutilig	manicio

Items	Qty
Main Unit	1
Power Cable	1
User Manual	1
Table 6	

98-V-B Rows Magnetic Heating Mantle

Items	Qty
Main Unit	1
Power Cable	1
User Manual	<u> </u>
Stirrer bar	Rows

Table 7

Please check the instrument and appendix with the packing list when you first open the instrument packing case. If you find there is something wrong with the instrument and the appendix, do contact the vendor or the producer.

4 Trial Runs

- · Make sure the required operating voltage and power supply voltage match.
- Ensure the socket must be earthed reliably.
- · Ensure the power be off
- Plug in the power cable, ensure the power be on and begin initializing.
- Add the medium into the vessel with a stirring bar if with the magnetic stirrer function.
- · Put the vessel on the work plate.
- Adjust the stirring speed and start stirring if with the magnetic stirrer function...
- · Observe the stirring bar and LCD display if with digital function.
- · Adjust the temperature and start heating.
- Observe the real temperature on LCD display if with the digital function.
- Stop the heating and stirring functions.

If these operations above are normal, the device is ready to operate. If these operations are not normal, the device may be damaged during transportation, please contact manufacture for technical support.

Table 5

5 Operating Modes


5.1 98-I-B Manual Heating Mantle and 98-IV-B Rows Manual Heating Mantle:

- Place the equipment on level worktable, and then put the container with liquid inside into heating mantle.
- Switch on the power accords with the machine, then power indicator light will be
 lighten; turn on the power of regulation knob, and turn the knob clockwise, then the
 working indicator light will be lighten; in the process of regulating, the light intensity
 changes according to different regulation position, and the temperature rises as well.

<u>∧</u>

Notes:

The power must accord with the machine.







5.2 98-II-B Magnetic Heating Mantle and 98-V-B Rows Magnetic Heating Mantle:


- Place the equipment on level worktable, and then put the container with liquid inside into heating mantle.
- Switch on the power accords with the machine, then power indicator light will be lighten; turn on the power of regulation knob, and turn the knob clockwise, then the working indicator light will be lighten; in the process of regulating, the light intensity changes according to different regulation position, and the temperature rises as well.
- · The same to adjust the magnetic stirring power.



Notes:

The power must accord with the machine.

Adjust the speech slowly; please adjust the speech when too high speech makes the stir breakaway Make sure the power line has safety distance from the heating mantle.



5.3 98-I-C Digital Heating Mantle:

- Place the equipment on level worktable, and then put the container with liquid inside into heating mantle.\
- Install the Sensor Rack with Stainless Rods on the heating mantle back holder.
- · Put the temperature sensor into the liquid.
- Turn the knob anticlockwise to the left place, Setting the inquiry temperature and the equipment will working slowly.
- If the temperature on the screen can't up to the setting temperature, turn the knob clockwise slowly, the real temperature on the screen will increase slowly.
- If the temperature still can't up to the setting temperature, do this process again.
- If the temperature higher than the setting temperature, turn the knob anticlockwise to make the temperature down



Notes:

The power must accord with the machine.

Make sure the power line has safety distance from the heating mantle.

When the machine meets fault please cut off the Electricity first.



5.4 98-II-C Digital Heating Mantle:

- Place the equipment on level worktable, and then put the container with liquid inside into heating mantle.
- Install the Sensor Rack with Stainless Rods on the heating mantle back holder.
- Put the temperature sensor into the liquid.
- Press the S keyboard comes into the setting condition. Adjust the temperature by press the up and down keyboard.
- Press the **S** keyboard again to finish the temperature setting and then the equipment comes into the working condition.

Notes:



The power must accord with the machine. Make sure the power line has safety G distance from the heating mantle.



When the machine meets fault please cut off the Electricity first.

5.5 98-III-B Magnetic and Digital Heating Mantle:

- Place the equipment on level worktable, and then put the container with liquid inside into heating mantle.\
- Install the Sensor Rack with Stainless Rods on the heating mantle back holder.
- Put the temperature sensor into the liquid.
- Press the S keyboard comes into the setting condition. Adjust the temperature by press the up and down keyboard.
- Press the **S** keyboard again to finish the temperature setting and then the equipment comes into the working condition.
- Turn the knob clockwise and anticlockwise to adjust the magnetic stirrer power.

Notes:



The power must accord with the machine. Make sure the power line has safety distance

from the heating mantle.

Adjust the speech slowly; please adjust the Speech when too high speech makes the stir breakaway

When the machine meets fault please cut off the Electricity first.



6 Faults

- · Instruments can't be power ON
 - Check whether the power cable is plugged
 - Check whether the fuse is broken or loose
- Temperature cannot reach set point or stirring can't be starts when adjust the control knob
 - Check whether the heating wire broke during transport
 - Check whether the controller broke during transport

If these faults are not resolved, please set the instruments to factory default setting, or take the unit to your technical service center, or contact with the manufacturer.

7 Maintenance and Cleaning

- Proper maintenance can keep instruments working in a good state and lengthen its lifetime.
- · Be careful not spray the cleanser into the instrument when cleaning.
- Unplug t he power line when cleaning.
- Only use cleanser that we advised as below:

Dyes	Isopropyl alcohol
Construction materials	Water containing tenside /
	Isopropyl alcohol
Cosmetics	water containing tenside /
	Isopropyl alcohol
Foodstuffs	Water containing tenside
Fuels	Water containing tenside

Table 9

8 Storage and transportation

· Keep it in dry and clean room with good ventilation and no corrosive gas

• prevent it from wetting by the rain and avoid violent collision in transportation.

9 Working condition

Ambient temperature: 5~40°C;

Ambient humidity: ≤90%;

Voltage: 220V±10%, 50/60Hz or 110V+/-10%, 50/60Hz

10 Main technical parameters

Model	Capacity	Voltage	Max.	Power		Exterior Size	Packing Size	N.W.	G.W.
model	(ml)	(V)	Temp.	(W)	Time	(mm)	(mm)	(KG)	(KG)
	50			80		Φ200x160	230x230x170	2	2.5
	100]		100					
	250]		150		W200X100	23022302170	2	2.5
	500]		250					
98-I-B	1000	1		350		Ф260x200	290x290x220	3.5	4
98-I-B	2000	1		450		Ф300x230	330x330x250	4	5
	3000	1		600		Ф300x250	330x330x270	6	7
	5000	1		800		Ф350x270	380x380x290x	7	8
	10000	1		1200	000 000 000 Continu 000 ous	Φ420x320	450x450x340	10	12
1	20000	1	420℃	2500		450x450x380	500x500x400	21	26
	50	220V,		80		Ф200x160	230x230x170	2.5	2.8
	100	50/60Hz		100					
	250			150					
	500	110V/60H		250					
00 11 0	1000	Z		350		Ф260x200	290x290x220	4	4.5
98-II-B	2000	1		450		Ф300х230	330x330x250	5	6
	3000	1		600	1	Ф300x250	330x330x270	7	8
	5000	1		800	0	Ф350x270	380x380x290x	9	10
	10000	1		1200		Φ420x320	450x450x340	12	14
	20000	1		2500		450x450x380	500x500x400	23	28
	50	1		80			230x230x170	2.5	2.8
0010	100	1		100					
98-I-C	250	1		150		Φ220x160			
	500	1		250					

Table 10

	1000	350	Ф260x220	290x290x220	5.5	6
1	2000	450	Ф300x230	330x330x250	6.5	7
I	3000	600	Ф300x250	300x330x270	7.5	8
	5000	800	Ф350x270	380x380x290	8.5	9.2
1	10000	1200	Φ420x320	450x450x340	9.8	12
	20000	2500	450x450x380	500x500x400	21	26
	50	80				
I	100	100		220-245-405	2.5	
1	250	150	Φ220x165	230x215x195	2.5	2.8
1	500	250				
00 11 0	1000	350	Ф270x220	280x280x300	5.5	6
98-II-C	2000	450	Ф330x230	345x290x350	6.5	7
	3000	600	Ф340x245	365x365x295	7.5	8
	5000	800	Ф350x250	390x390x310	8.5	9.2
	10000	1200	Φ425x320	470x470x380	9.8	12
	20000	2400	550x510x390	540x540x420	21	26
	100	100	Φ220x165	230x215x195	2.5	2.8
1	250	150				
l	500	250				
	1000	350	Ф 270x220	280x280x300	5.5	6
98-III-B	2000	450	Ф330x230	345x290x350	6.5	7
	3000	600	Ф340x245	365x365x295	7.5	8
1	5000	800	Ф350x250	390x390x310	8.5	9.2
1	10000	1200	Φ425x320	470x470x380	9.8	12
	20000	2400	550x510x390	540x540x420	21	26
98-IV-B	100	100x2	280x140x106	320x180x190	4	5
Two Rows	250	150x2	320x160x170	360x190x200	4	5

	500	250x2	360x180x180	400x220x210	4	
-	1000	350x2	420x210x200	460x250x230	7	-
	100	100x4	530x140x160	570x180x190	8	1
98-IV-B	250	150x4	610x160x170	650x190x200	8	1
Four Rows	500	250x4	690x180x180	730x220x210	8	1
11.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1	1000	350x4	810x210x200	850x250x230	14	1
	100	100x6	820x140x160	860x180x190	12	1
98-IV-B	250	150x6	940x160x170	980x190x200	12	1
Six Rows	500	250x6	1060x180x180	1100x220x210	12	1
27 40 04 300 0.700	1000	350x6	1240x210x200	1280x250x230	21	2
	100	100x2	280x140x106	320x180x190	4	Ę
98-V-B	250	150x2	320x160x170	360x190x200	4	Ę
Two Rows	500	250x2	360x180x180	400x220x210	4	Ę
	1000	350x2	420x210x200	460x250x230	7	8
	100	100x4	530x140x160	570x180x190	8	1
98-V-B	250	150x4	610x160x170	650x190x200	8	1
Four Rows	500	250x4	690x180x180	730x220x210	8	1
	1000	350x4	810x210x200	850x250x230	14	1
	100	100x6	820x140x160	860x180x190	12	1
98-V-B	250	150x6	940x160x170	980x190x200	12	1
Six Rows	500	250x6	1060x180x180	1100x220x210	12	1
	1000	350x6	1240x210x200	1280x250x230	21	2

Table 11